

VERIFICATION OF COMPLIANCE

- **Equipment : AC-DC Switching Power Supply**
Model No. : MFC15-xy, where: "x" can be 3.75 ~ 5.25, 6.75 ~ 9.45, 9 ~ 12.6, 11.25 ~ 15.75 or 18 ~ 25.2; "y" can be S, S-A2 or S-A2-DN
MZC20-xy, where: "x" can be 9 ~ 12.6 or 18 ~ 25.2;
"y" can be S, S-A2 or S-A2-DN
- Applicant : ARCH ELECTRONICS CORP**
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi
Dist., New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard**

EN 55011:2009+A1:2010 Class B Group 1 Equipment,

EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015

(IEC 61000-4-2 Edition 2.0 2008-12, IEC 61000-4-3 Edition 3.2 2010-04,

IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05,

IEC 61000-4-6 Edition 4.0 2013-10, IEC 61000-4-8 Edition 2.0 2009-09,

IEC 61000-4-11 Edition 2.0 2004-03).

The test was carried out on **Jun. 14, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

William Li
Supervisor