

VERIFICATION OF COMPLIANCE

- **Equipment** : AC-DC Switching Power Supply
- Model No.** : MQCS100-xS, MQCS100-xS-A2, MQF120O-xS, MQF120U-xS, MQF120E-xS
(where "x" can be 9 to 15 ; 13.5 to 16.5 ; 18 to 30 ; 36 to 54)
- Applicant** : ARCH ELECTRONICS CORP
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi Dist.,
New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment was **Passed** the test performed according to
European Standard EN 60601-1-2:2015

**(EN 55011:2009+A1:2010 Class B Group 1 Equipment,
EN 61000-3-2:2014, EN 61000-3-3:2013, IEC 61000-4-2 Edition 2.0 2008-12,
IEC 61000-4-3 Edition 3.2 2010-04, IEC 61000-4-4 Edition 3.0 2012-04,
IEC 61000-4-5 Edition 3.0 2014-05, IEC 61000-4-6 Edition 4.0 2013-10,
IEC 61000-4-8 Edition 2.0 2009-09, IEC 61000-4-11 Edition 2.0 2004-03)**

The test was carried out on **Dec. 04, 2017** at **SPORTON INTERNATIONAL INC. EMC &
Wireless Communications Laboratory.**

A handwritten signature in blue ink, appearing to read 'William Li', written over a horizontal line.

William Li

SPORTON INTERNATIONAL INC. EMC & Wireless Communications Laboratory

No. 52, Huaya 1st Rd., Guishan Dist., Taoyuan City, Taiwan (R.O.C.)