

# VERIFICATION OF COMPLIANCE

- **Equipment** : AC-DC Switching Power Supply
- Model No.** : MQF500O-xS, MQF500U-xS, MQF500E-xS  
where "x" can be 9 ~12,48 ; 11.25 ~ 15.6 ; 18 ~ 24,96 ;  
36 ~ 49,92 ; 39 ~ 54.08
- Applicant** : ARCH ELECTRONICS CORP  
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi Dist.,  
New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard**

**EN 55011:2009+A1:2010 Class B Group 1 Equipment,**

**EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015**

**(IEC 61000-4-2 Edition 2.0 2008-12, IEC 61000-4-3 Edition 3.2 2010-04,**

**IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05,**

**IEC 61000-4-6 Edition 4.0 2013-10, IEC 61000-4-8 Edition 2.0 2009-09,**

**IEC 61000-4-11 Edition 2.0 2004-03).**

The test was carried out on **Sep. 25, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

A handwritten signature in blue ink, appearing to read 'William Li', written over a horizontal line.

**William Li**  
**Supervisor**