

# VERIFICATION OF COMPLIANCE

- **Equipment** : AC-DC Switching Power Supply
- Model No.** : MSC60-xy, where: "x" can be 3.75 ~ 5.25, 6.75 ~ 9.45, 9 ~ 12.6, 11.25 ~ 15.75 or 18 ~ 25.2; "y" can be S, S-A2 or S-A2-DN
- Applicant** : ARCH ELECTRONICS CORP  
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi Dist., New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard**

**EN 55011:2009+A1:2010 Class B Group 1 Equipment,**

**EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015**

**(IEC 61000-4-2 Edition 2.0 2008-12, IEC 61000-4-3 Edition 3.2 2010-04,**

**IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05,**

**IEC 61000-4-6 Edition 4.0 2013-10, IEC 61000-4-8 Edition 2.0 2009-09,**

**IEC 61000-4-11 Edition 2.0 2004-03).**

The test was carried out on **Jun. 14, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

A handwritten signature in blue ink, appearing to read 'William Li', written over a horizontal line.

**William Li**  
**Supervisor**