

# VERIFICATION OF COMPLIANCE

- **Equipment : AC-DC Switching Power Supply**  
**Model No. : MTC-3.3S,MTC-5S,MTC-7.35S,MTC-9S,MTC-12S,MTC-15S,**  
**MTC-24S,MTC-5D,MTC-12D,MTC-15D,MTC-5S12D,MTC-5S15D**
- Applicant : ARCH ELECTRONICS CORP**  
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi  
Dist., New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard**

**EN 55011:2009+A1:2010 Class B Group 1 Equipment,**  
**EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015**  
(IEC 61000-4-2 Edition 2.0 2008-12,IEC 61000-4-3 Edition 3.2 2010-04,  
IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05,  
IEC 61000-4-6 Edition 4.0 2013-10,IEC 61000-4-8 Edition 2.0 2009-09,  
IEC 61000-4-11 Edition 2.0 2004-03).

The test was carried out on **Jun. 14, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

**William Li**  
**Supervisor**