


# VERIFICATION OF COMPLIANCE

- **Equipment : AC-DC Switching Power Supply**  
**Model No. : MTC30-xy, where: "x" can be 3.75 ~ 5.25, 9.0 ~ 12.6, 11.25 ~ 15.75 or 18.0 ~ 25.2; "y" can be S, S-A2 or S-A2-DN**  
**Applicant : ARCH ELECTRONICS CORP**  
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi  
Dist., New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard EN 55011:2009+A1:2010 Class B Group 1 Equipment, EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015 (IEC 61000-4-2 Edition 2.0 2008-12, IEC 61000-4-3 Edition 3.2 2010-04, IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05, IEC 61000-4-6 Edition 4.0 2013-10, IEC 61000-4-8 Edition 2.0 2009-09, IEC 61000-4-11 Edition 2.0 2004-03).**

The test was carried out on **Jun. 10, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

  
**Jack Deng**  
**Engineering Manager**