

# VERIFICATION OF COMPLIANCE

- **Equipment : AC-DC Switching Power Supply**  
**Model No. : MQC100-xS, MQC150-xS, MQF2400-xS,**  
**MQF240U-xS, MQF240E-xS**  
(where "x" can be 9 to 12.48 ; 11.25 to 15.6 ;  
18 to 24.96 ; 36 to 49.92)  
**Applicant : ARCH ELECTRONICS CORP**  
3F., No. 79, Sec. 1, Xintai 5th Rd., Xizhi Dist.,  
New Taipei City 22101, Taiwan

**I HEREBY****DECLARE THAT :**

The equipment is in accordance with the procedures are given in **EUROPEAN COUNCIL DIRECTIVE 93/42/EEC**. The equipment was **Passed** the test performed according to **European Standard**

**EN 55011:2009+A1:2010 Class B Group 1 Equipment,**

**EN 61000-3-2:2014, EN 61000-3-3:2013 and EN 60601-1-2:2015**

**(IEC 61000-4-2 Edition 2.0 2008-12, IEC 61000-4-3 Edition 3.2 2010-04,**

**IEC 61000-4-4 Edition 3.0 2012-04, IEC 61000-4-5 Edition 3.0 2014-05,**

**IEC 61000-4-6 Edition 4.0 2013-10, IEC 61000-4-8 Edition 2.0 2009-09,**

**IEC 61000-4-11 Edition 2.0 2004-03).**

The test was carried out on **Sep. 07, 2017** at **SPORTON INTERNATIONAL INC. LAB.**

A handwritten signature in blue ink, appearing to read 'William Li', is written over a horizontal line.

**William Li**  
**Supervisor**