

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply for building-in

Name and address of the applicant

Arch Electronics Corp.
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei
County 221, Taiwan

Name and address of the manufacturer

Arch Electronics Corp.
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei
County 221, Taiwan

Name and address of the factory

Arch Electronics Corp.
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei
County 221, Taiwan*Note: When more than one factory, please report on page 2* Additional Information on page 2

Ratings and principal characteristics

I/P: 100-240 Vac; 47-63 Hz; 0,8-0,4 A
O/P: See page 2.

Trademark (if any)



Customer's Testing Facility (CTF) Stage used

/

Model / Type Ref.

MSC-xyz
See page 2.

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed.

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005 + A1:2012

As shown in the Test Report Ref. No. which forms part of this Certificate

T223-0663/18, issued on 2019-02-25

This CB Test Certificate is issued by the National Certification Body

SIQ Ljubljana, Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

SIQ Ljubljana is accredited by Slovenian Accreditation with accreditation number CP-001 in the field of certification of products, processes and services.

Date: 2019-02-25

Signature: Bojan Pečavar

A handwritten signature in blue ink, appearing to read 'Bojan Pečavar', written over a horizontal line.

Ratings and principal characteristics:

I/P: 100-240 Vac; 47-63 Hz; 0,8-0,4 A

O/P:

Model Name	Output Ratings (output dc voltage / output current)
MSC-3.3S	3,3 V / 8 A
MSC-5S	5 V / 8 A
MSC-9S	9 V / 4,444 A
MSC-12S	12 V / 3,333 A
MSC-15S	15 V / 2,666 A
MSC-24S	24 V / 1,667 A
MSC-5D	+5 V / 4 A, -5 V / 4 A
MSC-12D	+12 V / 1,666 A, -12 V / 1,666 A
MSC-15D	+15 V / 1,333 A, -15 V / 1,333 A
MSC-5S12S	+5 V / 5 A, +12 V / 1,25 A
MSC-5S24S	+5 V / 5 A, +24 V / 0,625 A
MSC-5S12D	+5 V / 5 A, +12 V / 0,6 A, -12 V / 0,6 A
MSC-5S15D	+5 V / 5 A, +15 V / 0,5 A, -15 V / 0,5 A

Model / Type Ref.:

MSC-xyz

"x" can be 3,3; 5; 9; 12; 15 or 24

"y" can be S or D

"z" can be 12S, 24S, 12D, 15D or blank

Additional information (if necessary)

This CB test certificate substitutes previously issued CB test certificate No. SI-3638, issued on 2013-03-20, due to update of test report and standard update.

Date: 2019-02-25

Signature: Bojan Pečavar

