

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

AC/DC Switch Mode Power Supplies

Name and address of the applicant
Nom et adresse du demandeur

Arch Electronics Corp
7th Fl-1 79 Hsin Tai Wu Rd Sec 1 Hsin Chin New Taipei 221
Taiwan

Name and address of the manufacturer
Nom et adresse du fabricant

Arch Electronics Corp
7th Fl-1 79 Hsin Tai Wu Rd Sec 1 Hsin Chin New Taipei 221
Taiwan

Name and address of the factory
Nom et adresse de l'usine

ARCH ELECTRONICS CORP
3rd FL 79 SEC 1 HSIN TAI WU RD HSIN CHIN NEW TAIPEI
221
TAIWAN

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Input Rating: 100-240 VAC, 47-63 Hz, 2.0 A max.
Output Rating: See test report for details.

Trademark (if any)
Marque de fabrique (si elle existe)



Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

AQC100-XS-A2, AQC100-XS, AQF100C-XS, AQF100O-XS Y
See Page 2

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} page

Additionally evaluated to EN 60950-1:2006/ A11:2009/ A1:2010/
A12:2011/ A2:2013; National Differences specified in the CB
Test Report.

Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60950-1(ed.2), IEC 60950-1(ed.2);am1, IEC 60950-
1(ed.2);am2

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

1603051-CB issued on 2016-06-15

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-06-15

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-55340-UL

Model Details:

AQC100-XS-A2,AQC100-XS,AQF100C-XS,AQF100O-XS Y Where X=12 to 14.9, 15 to 23.9, 24 to 34.9 or 35 to 48;
Y= A or blank.

Additional information (if necessary)

Information complémentaire (si nécessaire)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-06-15

Signature:

Jan-Erik Storgaard