



Ref. Certif. No.

SI-3471

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE

## CERTIFICAT D'ESSAI OC

Product  
Produit

Power supply

Name and address of the applicant  
Nom et adresse du demandeurArch Electronics Corp.  
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County  
221, TaiwanName and address of the manufacturer  
Nom et adresse du fabricantArch Electronics Corp.  
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County  
221, TaiwanName and address of the factory  
Nom et adresse de l'usineArch Electronics Corp.  
3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County  
221, TaiwanNote: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page Additional Information on page 2Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

See page 2

Trademark (if any)  
Marque de fabrique (si elle existe)

ARCH

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.  
Ref. De typeAQF120X-YS; X can be O, U or E; Y can be 12S, 24S, 30S or  
48SAdditional information (if necessary may also be reported  
on page 2)Les informations complémentaires (si nécessaire, peuvent  
être indiqués sur la 2<sup>ème</sup> page Additional Information on page 2A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60950-1:2005 (2nd Ed.)+A1:2009

As shown in the Test Report Ref. No. which forms part of  
this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

T223-0322/12

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Slovenski inštitut za kakovost in meroslovje  
Slovenian Institute of Quality and Metrology  
Tržaška c. 2, SI-1000 Ljubljana, Slovenia  
Product Certification Body is accredited by Slovenian Accreditation, Reg. No.: CP-001

Date: 2012-09-14

Signature: Alja Pregl



## Ratings and principal characteristics:

Input: 100-240 Vac, 2 Amax, 47-63 Hz

Output: Model AQF120X-12S: 12Vdc, 10A;  
Model AQF120X-24S: 24Vdc, 5A  
Model AQF120O-30S: 30Vdc, 4A  
Model AQF120X-48S: 48Vdc, 2,5A

Max. output power: 120W;

Max. output power with 18CFM forced air cooling: 180W;

**Additional information (if necessary)****Information complémentaire (si nécessaire)**

This CB test certificate substitutes previously issued CB test certificate No. SI-2805, dated 2010-12-24. New CB test certificate issued due to new test report.

Date: 2012-09-14

Signature: Alja Pregl

